N°575 SPECTRO-3 series

Instruments

1. Color differentiation of metal rings

Metal rings should be color differentiated. For this purpose a color sensor type SPECTRO-3-5-DIF-JR is used. At this, the distance to the object is approximately 7mm and the detecting range at this distance is around 5mm in diameter. The metal rings can be proper differentiated as shown in the screen shots.











